





This is to certify that:

Great Eco Glove Sdn Bhd Lot 1675, Jalan Makmur Batu 28, Ijok Bestari Jaya Selangor 45600 Malaysia



CE 735448

In respect of:

Powder Free Nitrile Examination Gloves

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Drs. Dave Hagenaars, Managing Director

First Issued: 2020-11-03 Latest Issue: 2020-11-03 Effective Date: 2020-11-03 Expiry Date: 2025-11-03

Page: 1 of 4



...making excellence a habit."

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated <u>online</u>.

No. CE 735448

Product Specification	
Model – Powder Free Nitrile Examination Gloves	
Available sizes:	7 (S), 8 (M), 9 (L)
Colour:	Blue
Classification:	Protective glove for use against chemical and micro-organism hazards.
Description:	Powder Free Nitrile Disposable Protective Gloves manufactured from 100% nitrile synthetic rubber latex (Acylonitrile-butadiene) not containing natural rubber latex.
	The main features of this protective glove are:
	- Micro-organisms penetration resistance
	- Chemical permeation resistance
	- Powder free
	- Ambidextrous
Models in range:	Nitrile gloves.
Category:	Category III – complex
Applicable Standards:	General requirements for gloves to EN ISO 21420:2020 (not harmonised yet) Protective gloves again chemicals & micro-organisms to EN ISO 374-1:2016 Resistance to penetration to EN 374-2:2014 Determination of resistance to permeation by chemicals tested to EN 16523-1:2015 Resistance to penetration by blood borne tested to EN ISO 374-5:2016 Resistance to degradation by chemicals to EN 374-4:2013 Resistance of clothing materials to penetration by blood-borne pathogens to ISO 16604:2014

First Issued: 2020-11-03 Latest Issue: 2020-11-03 Effective Date: 2020-11-03 Expiry Date: 2025-11-03

Page: 2 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated <u>online</u>.

No. CE 735448

Product Specification (Continued)

Model – Powder Free Nitrile Examination Gloves 9NPF

Performance:

General requirements for gloves to EN ISO 21420:2020

pH Value: Pass PAH (Innocuousness): Pass Size: 7, 8, 9 Dexterity Level 5

Resistance to penetration to EN 374-2:2014 Pass

Resistance to chemical permeation to EN ISO 374-1:2016 Test method EN 16523-1:2015

40% Sodium hydroxide (K) - level 6 37% Formaldehyde (T) - level 2

Protection against micro-organism risks to EN ISO 374-5:2016

Bacteria and fungi (Test method EN 374-2:2014) Pass Viruses (Test Method ISO 16604:2004) Pass

Resistance to degradation by chemicals to EN 374-4:2013

40% Sodium hydroxide (K) -5.6% 37% Formaldehyde (T) 14.3%

First Issued: 2020-11-03 Latest Issue: 2020-11-03 Effective Date: 2020-11-03 Expiry Date: 2025-11-03

Page: 3 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.

To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

No. CE 735448

Certificate Administration Details

Technical File Reference: GEG-TF-N-PPE-001, Rev:00.

Certificate Amendment Record

Issue date **Comments** November 2020 First issue.

BSI Internal report No. 2797:20:3281103

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module C2) as referenced on BSI issued Certificate CE 738460.

First Issued: 2020-11-03 Latest Issue: 2020-11-03 Effective Date: 2020-11-03 Expiry Date: 2025-11-03

Page: 4 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.